

SKC Rotameters



Standard Rotameters

Description	Part Number
Rotameter 0.02-0.25 L/min Accuracy $\pm 2.5\%$	393-002025
Rotameter 0.3-3.4 L/min Accuracy $\pm 2.5\%$ VDI/VDE 3513-2	393-0334
Rotameter 0.6-5 L/min Accuracy $\pm 2.5\%$ VDI/VDE 3513-2	393-0650
Rotameter 1.0-13.0 L/min Accuracy $\pm 2.5\%$ VDI/VDE 3513-2	393-1130
Rotameter 2.0-26.0 L/min Accuracy $\pm 2.5\%$ VDI/VDE 3513-2	393-2260

Field Rotameters

Description	Part Number
Description	Part Number
Field Rotameter 50-500 ml/min (50 mm scale), Accuracy 5%	320-2A05
Field Rotameter 100-1000 ml/min (100 mm scale), Accuracy 3%	320-4A1
Field Rotameter 400-5000 ml/min (100 mm scale), Accuracy 3%	320-4A5
Field Rotameter 2-20 L/min (100 mm scale), Accuracy 3%	320-4A20L
Field Rotameter 3-30 L/min (100 mm scale), Accuracy 3%	320-530
Field Rotameter 3-30 L/min (100 mm scale)*, Accuracy 3%	320-100
Field Rotameter 4-50 L/min (100 mm scale), Accuracy 3%	320-440

* Fittings adapted for QuickTake 30 sample pump

Meetapparatuur voor professionals

